

응용 자료

Analysis of Cannabinoids in Cannabis Plant Materials and Edible Products Using UltraPerformance Liquid Chromatography (UPLC) with PDA and Mass Detection

Kim Van Tran, Marian Twohig, Christopher J. Hudalla

Waters Corporation, ProVerde Laboratories



Abstract

The legalization of cannabis for both medicinal purposes and adult use continues to advance. Consequently, there is a need for simple and reliable analytical methods for the analysis of plant material and cannabis-derived products to support label claims and to ensure quality and safety for consumers.

A robust analytical method has been developed to detect and quantify 18 cannabinoids in cannabis flower, hemp, and edible products using UPLC-PDA and mass detection. Quantitative data were acquired using an ACQUITY PDA Detector combined with the ACQUITY QDa Mass Detector, a single quadrupole mass detector, and controlled by Empower Chromatography Data Software. A Waters ACQUITY UPLC H-Class-PDA-QDa System combined with a CORTECS C₁₈ analytical column enabled the separation of 18 cannabinoids in less than 10 minutes. The lowest level used in the calibration curve was 0.4 mg/mL for cannabinoids detected using QDa mass detection, and the lowest level used in the calibration curve was around 3.125 µg/mL for cannabinoids detected using the PDA. The ACQUITY QDa provided more accurate detection for cannabinoids at low levels compared to the PDA. The results show that this method is suitable for the analysis and accurate detection of cannabinoids across a wide range of matrices including cannabis flower, hemp, concentrates, and edible products.

Benefits

- The ACQUITY UPLC-PDA-QDa System is a highly reliable and robust instrument that can be used for routine analysis of cannabis flower, hemp, concentrates, and edible products
- Enhanced confidence in peak identity confirmation is achieved using both PDA and mass detection
- Empower Chromatography Data Software manages all raw data, methods, and results by permanently linking electronic records together

Introduction

The legalization of cannabis for both medicinal purposes and adult use continues to advance. As more products are developed and enter the market, the need for simple, reliable, and fast analytical methods to support label claims and ensure quality and safety for consumers grows in importance. The cannabis plant (*Cannabis sativa*) is a complex natural product that produces hundreds of different cannabinoids, but most

laboratories focus their analyses on just five major compounds: Δ^9 -tetrahydrocannabinol (Δ^9 -THC), Δ^9 -tetrahydrocannabinolic acid (THCA), cannabidiolic acid (CBDA), cannabidiol (CBD), and cannabinol (CBN). Recently, other minor cannabinoids have shown beneficial medicinal effects, therefore increasing the importance of separating and identifying them alongside the major compounds. Minimizing peak co-elutions of minor and major cannabinoids in a reasonable timeframe, whilst providing accurate quantitative results, is desirable. This can be particularly challenging in edible products, as cannabinoids have been added to a wide range of foods, including those with high sugar and high fat, like chocolate and gummies. Chemically complex samples require effective sample preparation to ensure successful analysis.¹ UltraPerformance Liquid Chromatography (UPLC) in combination with UV detection enables potency determination in complex matrices as it can identify and quantify structurally similar major and minor cannabinoids and their different forms allowing the determination of the correct blend of compounds to provide the most efficacious results. The addition of mass detection can provide both increased specificity and sensitivity. In this study, an ACQUITY UPLC H-Class System with both the ACQUITY PDA and the ACQUITY QDa Mass Detector was used for the analysis of 18 cannabinoids in cannabis flower, hemp, and edible products, including gummies, chocolates, and beverages.

Experimental

Materials and Reagents

Standard Compounds

Cannabinoid standards were obtained from Cayman Chemical (Ann Arbor, Michigan) and Cerilliant/Sigma-Aldrich (St. Louis, MO).

Reagents

LC-MS grade solvents for sample extractions and LC mobile phases were obtained from Honeywell-Burdick and Jackson (Muskegon, MI). The formic acid was obtained from Sigma-Aldrich (St. Louis, MO).

Sample Preparation

Cannabis and derived samples were obtained from local sources (Massachusetts). Sample preparation varied depending on matrix.

Hemp and Cannabis Flower

Cannabis plant material (0.5 g) was weighed into 50 mL centrifuge tubes, and an aliquot of acetonitrile (20 mL) was added along with two stainless steel balls. The samples were processed with a Geno/Grinder® (SPEX, Metuchen, NJ) for 3 minutes at 1500 RPM (Figure 1a). The tubes were sonicated for 20 minutes and centrifuged for 5 minutes at 3000 RCF. After sonication, the samples were centrifuged, filtered through a 0.2 µm PTFE filter. Samples were analyzed in three replicates at two concentration levels at 80x and 2000x dilutions. Some of the main cannabinoids such as THCA and CBDA in the samples are significantly higher and require dilution to bring them into the calibration range.

Edibles

An aliquot (1 g) of freeze-ground or homogenized samples was weighed out and added to 10 mL water into a 50 mL centrifuge tube. The tube was vortexed for 1 min and sonicated for 20 min. Acetonitrile (10 mL) was added and the tube was vortexed for 1 min. The CEN QuEChERS salts (p/n 186006813) were added to the tube and shaken for 1 min followed by centrifugation for 5 minutes at 3000 RCF. The top acetonitrile layer can be injected directly or diluted with acetonitrile depending on the concentration of the samples.

The Freezer/Mill® was used for grinding gummies and sticky materials; however, embrittled gummy material will become sticky if left at room temperature for extended periods of time (Figure 1b).



Figure 1. Geno/Grinder® used for solvent extractions of solid material (panel A). The stainless steel balls facilitate comminution and can be used for shaking during the QuEChERS extraction. The Freezer/Mill® is used to pulverize gummy and sticky materials (panel B).

Cannabinoid-infused Beverages

An aliquot of the infused beverage (10 mL) was added to 10 mL of acetonitrile. After the sample was vortexed for 1 minute, CEN QuEChERS salts were added, and the sample was shaken for 1 minute and centrifuged for five minutes at 3000 RCF. The top acetonitrile layer can be injected directly or diluted with acetonitrile depending on the concentration of the samples.

Recovery of Cannabinoids from Gummy Matrix using QuEChERS Sample Preparation: (n=3)

Preparation of Pre-spiked Samples:

An aliquot (1000 µL) of a spiking solution containing 18 authentic cannabinoid standards at a concentration of 50 µg/mL was spiked into 1 g of homogenized gummy#941 (predetermined to contain 0.119 % CBD) for a final added concentration of 0.005% in gummy. Water (10 mL) was added and the sample was vortexed and sonicated for 20 min. Acetonitrile (9 mL) was added, the mixture was vortexed and the CEN QuEChERS salts were added. The tubes were shaken for 1 minute and centrifuged for 5 minutes. In the final step, 100 µL of acetonitrile was added to 1 mL acetonitrile extract from the QuEChERS step to make a final volume of 1.1 mL with posts-spiked samples.

Preparation of Post-spiked Samples:

Homogenized gummy#941 (1 g) (predetermined to contain 0.119 % CBD) was added into 10 mL water. The sample was vortexed and sonicated for 20 minutes, followed by the addition of 10 mL of acetonitrile. Next, CEN QuEChERS salts were added. The sample was then shaken for 1 minute and centrifuged for 5 minutes. The final acetonitrile extract (1 mL) was spiked with a mixture containing 18 cannabinoid standards (100 µL of 50 µg/mL).

The concentration of the samples was calculated against a standard curve prepared in solvent and the % recovery of the cannabinoids was determined.

LC Conditions

LC system:	ACQUITY UPLC H-Class
Detection:	PDA single wavelength @ 228 nm, 253 nm PDA Spectrum 210–400 nm at 4.8 nm resolution
Vials:	Certified vial (p/n: 186005668CV)
Filter:	Syringe Filter (p/n: WAT200556)
Column(s):	CORTECS C ₁₈ , 1.6 µm, 2.1 mm x 150 mm (p/n: 186007096)

Column temp.:	29 °C
Sample temp.:	5 °C
Injection volume:	1 µL
Flow rate:	0.45 mL/min
Mobile phase A:	20 mM Ammonium formate pH 2.92
Mobile phase B:	Acetonitrile
Weak wash:	90:10, water:methanol
Strong wash:	5:95, water:acetonitrile
Seal wash:	90:10, water:methanol

Gradient Table

Time (min)	Flow (mL/min)	%A	%B	Curve
0	0.45	24	76	6
6.4	0.45	24	76	6
6.5	0.45	1	99	6
8.0	0.45	1	99	6
8.1	0.45	24	76	1
10.0	0.45	24	76	1

MS Conditions

MS system:	ACQUITY QDa
Ionization mode:	Positive and negative ion electrospray (ESI+/ESI-)

Acquisition mass range:	100–600 Da SIR ESI+ and SIR ESI-
Capillary voltage:	1.5 kV (+), 0.8 kV (-)
Cone voltage:	10 kV (+), 15 kV (-)
Source temp.:	150 °C
Probe temp.:	450 °C

Data Management

Chromatography software:	Empower Chromatography Data Software (CDS)
--------------------------	--

Results and Discussion

PDA and Mass Detection

The analysis of 18 cannabinoids listed in Table 1 was performed using UPLC with PDA and mass detection. The ACQUITY QDa is a robust mass detector designed to be integrated into chromatography workflows for applications that would benefit from mass spectral information. In complex matrices, mass detection can increase confidence in peak identification and can enable lower detection limits.

MS data were collected in full scan ESI positive and negative mode in combination with more specific Selected Ion Recording experiments (SIR). The PDA was set up to monitor individual wavelengths (228 nm and 253 nm) and collect full PDA spectra from 210 to 400 nm. Retention times, PDA data, and mass spectral data together were used to identify detected cannabinoids in the samples (Figure 2). Figure 3 shows the instruments, software, and the analytical workflow that was used to analyze the cannabinoids.

	RT	Name	MWT	Chemical formula	CAS numbers
1	1.28	Cannabidivarin acid or Cannabidivarinic acid (CBDVA)	330	C ₂₀ H ₂₆ O ₄	31932-13-5
2	1.49	Cannabidivarin (CBDV)	286	C ₁₉ H ₂₆ O ₂	24274-48-4
3	1.82	Cannabidiol acid (CBDA)	358	C ₂₂ H ₃₀ O ₄	1244-58-2
4	1.92	Cannabigerolic acid (CBGA)	360	C ₂₂ H ₃₂ O ₄	25555-57-1
5	2.11	Cannabigerol (CBG)	316	C ₂₁ H ₃₂ O ₂	25654-31-3
6	2.25	Cannabidiol (CBD)	314	C ₂₁ H ₃₀ O ₂	13956-29-1
7	2.52	Tetrahydrocannabivarin (THCV)	286	C ₁₉ H ₂₆ O ₂	31262-37-0
8	3.04	Tetrahydrocannabivarinic acid (THCVA)	330	C ₂₀ H ₂₆ O ₄	39986-26-0
9	3.56	Cannabinol (CBN)	310	C ₂₁ H ₂₆ O ₂	521-35-7
10	3.70	Cannabinolic acid (CBNA)	354	C ₂₂ H ₂₆ O ₄	2808-39-1
11	4.37	exo-THC	314	C ₂₁ H ₃₀ O ₂	27179-28-8
12	4.61	Δ^9 -Tetrahydrocannabinol (Δ^9 -THC)	314	C ₂₁ H ₃₀ O ₂	1972-08-3
13	4.79	Δ^8 -Tetrahydrocannabinol (Δ^8 -THC)	314	C ₂₁ H ₃₀ O ₂	5957-75-5
14	5.61	Cannabicyclol (CBL)	314	C ₂₁ H ₃₀ O ₂	21366-63-2
15	5.70	Δ^9 -Tetrahydrocannabinolic acid A (THCA)	358	C ₂₂ H ₃₀ O ₄	23978-85-0
16	6.02	Cannabichromene (CBC)	314	C ₂₁ H ₃₀ O ₂	20675-51-8
17	6.42	Cannabichromenic acid (CBCA)	358	C ₂₂ H ₃₀ O ₄	185505-15-1
18	7.18	Cannabicyclolic acid (CBLA)	358	C ₂₂ H ₃₀ O ₄	40524-99-0

Table 1. List of 18 cannabinoids, observed retention times, chemical formulae, and CAS numbers.

Combining chromatographic, UV and mass data in a single place in the software can ease the burden of data interpretation. The Empower Mass Analysis window (Figure 2) provides a single location to associate chromatographic peaks from all detectors used in the analysis with their corresponding spectra which included the UV chromatogram and spectra are displayed along with the total ion chromatogram (TIC) and mass spectra and extracted ion chromatograms (XIC). Spectra from the detected peaks are time-aligned and displayed in a window above the chromatograms facilitating rapid data review. A chromatogram showing the separation of 18 cannabinoids is displayed in Figure 2, with a total method cycle time of 10 minutes.




Figure 2. Empower Mass Analysis window, shown in a single window. UV and MS spectra along with PDA and total ion chromatograms (TIC) and extracted ion chromatograms (XIC) can be viewed in a single window.




Figure 3. Analytical tools used for analysis of 18 cannabinoids in cannabis, hemp, edibles, and beverages.

Quantitation of Cannabinoids

Multi-point calibration curves for 18 cannabinoids prepared via serial dilution in acetonitrile were generated and showed good linearity for both PDA and mass detection ($R^2 > 0.99$). The calibration curves ranged from 3.1 to 50 $\mu\text{g}/\text{mL}$ for UV data at 228 nm and 0.4 to 50 $\mu\text{g}/\text{mL}$ for the ACQUITY QDa. The CBDV calibration curve derived from ACQUITY QDa data had an R^2 of 0.9878 due to the saturation of the detector at 50 $\mu\text{g}/\text{mL}$ (Figure 4). A calibration range of 0.4–25 $\mu\text{g}/\text{mL}$ can be used for some cannabinoids that have a high response with ACQUITY QDa detection and need further dilution to fit the linear curve.




Figure 4. Calibration curves for 18 cannabinoids at 228 nm and individual SIR channels from 0.1–50 µg/mL.

Figure 5 shows a chromatogram of the lowest calibration point used in the study for cannabinoids detected using the PDA at 228 nm and QDa SIR channels. The UV chromatogram shows the detection of the cannabinoids at 3.125 µg/mL with a 1 µL injection. The lowest calibration point for mass detection shown is 0.4 µg/mL.




Figure 5. Chromatograms of the lowest calibration standard tested at 228 nm (3.15 µg/mL) and for the individual SIR experiments (0.4 µg/mL).

Using a targeted MS experiment (SIR) allows improved sensitivity and specificity, enabling lower limits of detection in complex matrices. Representative chromatograms from the analysis of low Δ^9 -THC variety and high Δ^9 -THC variety cannabis flowers are shown in Figures 6 and 7. In Figure 6, Empower CDS automatically identified and labelled CBD and CBDA in the UV chromatogram based upon retention time. The specificity obtained using targeted analysis is also apparent, increasing the confidence in identified components.

Negative ion ESI using SIR targeting m/z 357 was used to detect the acidic cannabinoids, like CBDA. In addition to the target SIR and extracted wavelength chromatograms, full scan MS data and PDA spectra from 210–400 nm were simultaneously recorded allowing the MS and UV spectra to be viewed and thus further increasing confidence in the identified peaks.




Figure 6. Chromatograms confirming peak identity of a low Δ^9 -THC variety flower sample analyzed using UV spectrum from 210–400 nm with both positive and negative MS TIC and UV at 228 nm and SIR of m/z 315 and m/z 357.

Similarly, in Figure 7, the Empower CDS identified and tagged Δ^9 -THC and THCA in the UV chromatogram based on the retention times. The SIR channels of m/z 315 and m/z 357 show the SIR chromatograms for Δ^9 -THC and THCA.




Figure 7. Chromatograms from the analysis of a high Δ^9 -THC variety flower sample analyzed using UV at 228 nm and SIR of m/z 315 and m/z 357. MS analysis increases the confidence in the identification of Δ^9 -THC in the sample.


An Empower report showing the results from the analysis of a low Δ^9 -THC variety cannabis sample is shown in Figure 8. The calculated concentrations are displayed in the table in both $\mu\text{g/mL}$ and also weight percent (wt%). Custom calculations can be designed to automatically calculate and report total THC and total CBD, as well as CBD/THC ratio. This reduces the need to perform the calculations separately and keeps the data in a single software environment. The Empower report is completely customizable to report relevant data.

PeakName	RT	Area	%Area	Amount ug	Wt%	
1	CEBDA	1.272	45157	5.64	0.20	0.396
2	CEBD	1.472	5570	0.70	0.10	0.192
3	CBDA	1.814	520047	64.97	5.31	10.617
4	CBGA	1.915	11453	1.43	0.11	0.212
5	CBD	2.236	74203	9.27	1.35	2.700
6	exo-THC	4.476	10188	1.27	0.17	0.344
7	delta 9 THC	4.575	8818	1.10	0.25	0.494
8	THCA	5.760	10871	1.36	0.06	0.119
9	CBC	5.974	13440	1.68	0.05	0.110
10	CBCA	6.377	14165	1.77	0.24	0.477
11	CBLA	6.836	4279	0.53	0.07	0.142

Empower 3

Cannabinoid Analysis

SAMPLE INFORMATION			
Sample Name:	hemp 7.25 0.5g/ 20ml 201980 2	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	Edibles hemp17 NDV20
Val:	1E8	Acq Method/Kit:	20141822 NDV200 29C resMS
Injection#:	1	Processing Method:	FDAC28_19NDV200
Injection Volume:	100 uL	Channel Name:	FDACR2 228nm @ 8m
Run Time:	80 Minutes	Proc. Chnl. Descr.:	FDACR2 228nm @ 8m
Date Acquired:	11/17/2020 5:23:40 PM EST		
Date Processed:	11/18/2020 12:02:50 PM EST		



Total CBD
CBD + 0.877 (CBDA)

Total THC
THC + 0.877 (THCA)

Total THC Wt% = 0.589
Total CBD Wt% = 12.010
CEC:THC Ratio = 20.065

Reported by: User: System
 Report Method: Cannabinoid Analysis
 Report Method ID: 1782_1785/1
 Page: 1 of 1

Project Name: Cannabis Analysis/Cannabis Potency Testing
 Date Printed: 11/17/2021
 1143.37/FMUS/Estern

Figure 8. Empower report showing UV chromatogram at 228 nm resulting from the analysis of a hemp sample. Custom calculations like total THC and total CBD are also shown in the report.

Comparison Between Detected Cannabinoids in Cannabis, Hemp, Edibles, and Cannabinoid-infused Beverages and Associated Label Claims

The quantitative results generated from the analysis of 18 cannabinoids in cannabis flower and hemp are shown in Table 2 and were calculated by wt%. Samples were from different manufacturers that contained various types and concentrations of cannabinoids. Samples with high THCA or CBDA were diluted 2000x to bring the samples within the calibration range. Minor cannabinoids at lower concentrations required less dilution (80x dilution). Most of the detected amounts were between 53% to 121% compared to values on the label claim. Some cannabinoids had high %RSD due to the low signal in the detector.

Sample name	Flower 065	Label 065	Flower 723	Label 723	Flower 725	Label 725	Flower 407	Label 407	Flower 409	Label 409	Flower 414	Label 414	
	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	
1	CBDVA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	
2	CBDV	nd	nd	0.043 (13) 108	0.04	nd	nd	nd	nd	nd	nd	nd	
3	CBDA	12.858 (2) 89	14.430	7.749 (1) 83	9.372	11.997 (22) 85	14.108	0.027 (10) 95	0.028	0.036 (3) 89	0.041	0.021 (12) 67	0.031
4	CBGA	0.217 (5) 53	0.409	0.181 (3) 62	0.294	0.323 (2) 80	0.402	0.170 (7) 95	0.178	0.417 (2) 83	0.505	0.113 (11) 73	0.154
5	CBG	nd	0.039	nd	nd	nd	nd	0.089 (13) 112	0.079	0.084 (6) 88	0.095	0.076 (8) 86	0.089
6	CBD	1.969 (1) 141	1.398	4.511 (2) 100	4.479	3.144 (1) 102	3.089	nd	nd	nd	nd	nd	nd
7	THCV	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
8	THCVA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
9	CBN	nd	nd	nd	0.008	nd	nd	nd	0.002	nd	nd	nd	0.006
10	CBNA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
11	exo-THC	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
12	Δ ⁹ -THC	0.198 (9) 121	0.163	0.392 (1) 103	0.381	0.254 (3.59) 91	0.278	0.521 (9) 88	0.593	0.600 (4) 71	0.856	0.640 (14) 68	0.945
13	Δ ⁸ -THC	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
14	CBL	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
15	THCA	0.278 (3) 75	0.374	0.077 (13) 51	0.152	0.248 (1) 76	0.326	13.027 (8) 94	13.816	16.522 (5) 84	19.789	10.282 (9) 74	13.961
16	CBC	0.170 (5) 173	0.103	0.234 (2) 107	0.217	0.285 (2) 97	0.291	nd	nd	nd	nd	nd	nd
17	CBCA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
18	CBLA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd

Table 2. Quantitative results from the analysis of cannabis and hemp flower.

%wt¹: % weight detected

%RSD: Relative Standard Deviation (n=3)

% Label: % Detected amount compared to amount Label claimed

%wt²: % weight on Label

The quantitative results for the cannabis edibles and cannabinoid infused beverages are shown in tables 3a and 3b respectively. The cannabinoids detected in edibles and infused beverages were calculated by %. Most of the detected amounts of cannabinoids ranged from 60–125% of label claim, except for gummy #868, where the detected value was 36% of the label claim.

	Name	Chocolate 154	Label 154	Chocolate 515	Label 515	Chocolate 695	Label 695	Gummy 868	Label 868	Gummy 632	Label 632	Gummy 941	Label 941
		%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²
1	CBDVA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
2	CBDV	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
3	CBDA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
4	CBGA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
5	CBG	0.003 (2) 103	0.003	nd	nd	0.004 (14) 105	0.004	nd	0.002	nd	nd	nd	nd
6	CBD	nd	0.001	nd	nd	nd	0.002	0.110 (6) 36	0.313	0.275 (9) 111	0.246	0.104 (5) 87	0.119
7	THCV	nd	nd	nd	nd	nd	0.001	nd	nd	nd	nd	nd	nd
8	THCVA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
9	CBN	nd	0.003	0.001 (5) 91	0.001	0.005 (17) 125	0.004	nd	0.002	nd	nd	nd	nd
10	CBNA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
11	exo-THC	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
12	Δ^9 -THC	0.367 (11) 107	0.343	0.124 (2) 107	0.115	0.489 (3) 111	0.440	nd	nd	nd	nd	nd	nd
13	Δ^8 -THC	0.010 (22) 83	0.012	0.003 (28) (0)* 71	0.004	0.010 (13) 60	0.017	nd	nd	nd	nd	nd	nd
14	CBL	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
15	THCA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
16	CBC	0.003 (20) 81	0.004	nd	0.001	0.004 (9) 78	0.005	0.009 (15) 86	0.011	nd	nd	nd	nd
17	CBCA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd
18	CBLA	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd	nd

Table 3a. Quantitative results from the analysis of edible products.

%wt¹: % weight detected

%RSD: Relative Standard Deviation (n=3) *%RSD by QDa

% Label: % Detected amount compared to amount Label claimed

%wt²: % weight on Label

Name	Coffee 705	Label 705	Strawberry 648	Label 648	Lemon 705	Label 705
	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²	%wt ¹ (%RSD) %Label	%wt ²
CBD	0.004 (1) 100	0.004	0.006 (9) 100	0.006	0.003 (1) 100	0.003

Table 3b. Quantitative results from the analysis of infused drinks. Only CBD was detected in these samples.

%wt¹: % weight detected

%RSD: Relative Standard Deviation (n=3)

% Label: % Detected amount compared to amount listed on label

%wt²: % weight on label

Figure 9 shows representative chromatograms from the analysis of cannabinoids in chocolate sample #515. Δ^9 -THC showed a high UV and QDa response. The signal for Δ^8 -THC was less for UV compared to QDa, resulting in a %RSD of 28% when analyzed by UV and 0% when analyzed by QDa. MS analysis increased the confidence of identification of Δ^9 -THC and Δ^8 -THC in the chocolate sample.




Figure 9. Chromatograms showing Δ^9 -THC and Δ^8 -THC in chocolate samples detected with UV at 228 nm and SIR of m/z 315.

Recoveries of Cannabinoids in Edibles and Cannabinoid-infused Beverages using QuEChERS Sample Preparation Method

Quantitative results for the % recoveries of spiked cannabinoids in edibles are shown in Figure 10. Recoveries were calculated by comparing the wt% for samples spiked prior to QuEChERS extraction (pre-spiked samples) with the wt% for samples spiked after QuEChERS extraction (post-spiked samples). The concentration of the spiked gummy at the 0.005% level (5 mg/mL in the samples) is close to the detection limit of the calibration curve in the UV analysis (3.15 $\mu\text{g/mL}$). This could affect recoveries of cannabinoids in some of the matrices at this level due to matrix effects. The recoveries of cannabinoids analyzed with the PDA ranged from 60% to 125%. The recoveries for cannabinoids analyzed with an ACQUITY QDa Mass Detector ranged from 94% to 106%, indicating improved accuracy over UV detection methods.




Figure 10. Recoveries for 18 cannabinoids in gummy at 0.005% level by using CEN QuEChERS analyzed by PDA and QDa (n=3).

A comparison of the detector response obtained for the cannabinoids using the PDA at 228 nm and the SIR channels is shown in Figure 11. The detector response for the cannabinoid mixture spiked at a level of 5 μg/mL in the gummy matrix is higher using the MS which can aid in quantifying cannabinoids at lower levels in cannabinoid-infused products.




Figure 11. UV chromatogram at 228 nm and superimposed SIR channels for the detection of 18 cannabinoids spiked into the gummy matrix at 5 µg/g (0.005%).

Conclusion

The Waters ACQUITY UPLC H-CLASS-PDA-QDa System combined with the CORTECS C₁₈ Column enabled effective separation of 18 cannabinoids in under 10 minutes.

The ACQUITY QDa Mass Detector provided orthogonal detection to PDA for the confirmation of peak identities and quantitation of cannabinoids at low levels. The lowest level used in the calibration curve was 0.4 µg/mL for cannabinoids using mass detection, and the lowest level used in the calibration curve was around 3.125 µg/mL for cannabinoids using PDA.

The Empower Mass Analysis window provides a single location to associate the chromatograms and spectra from all detectors used in the analysis. The consolidation of this information in one place makes data review and interpretation easier to manage.

Empower CDS has numerous features which help in analysis of the data including tailored calculations allow relevant information to be derived quickly. Performing the calculations within the software maintains integrity and aids in record keeping.

QuEChERS extraction is effective for extraction of cannabinoids in edibles and drinks with high recoveries.

This analytical workflow is suitable for the analysis of cannabinoids across a wide range of matrices including cannabis flower, hemp, concentrates and edibles, and beverages.

References

1. Aubin AJ, Layton C, Helmueller C. Separation of 16 Cannabinoids in Cannabis Flower and Extracts using a Reversed-Phase Isocratic HPLC Method. Waters Application Note. [720006426EN <https://www.waters.com/webassets/cms/library/docs/720006426en.pdf>](#) . (2018).
2. Dubrow GA, Pawar RS, Srigley C, Sam JF, Talavera C, Parker CH, Noonan GO. A Survey of Cannabinoids and Toxic Elements in Hemp-Derived Products from the United States Marketplace. *J. Food Compos. Anal.* , 97, 2021, 103800.

Featured Products

[ACQUITY UPLC H-Class PLUS System <https://www.waters.com/10138533>](https://www.waters.com/10138533)

[ACQUITY QDa Mass Detector <https://www.waters.com/134761404>](https://www.waters.com/134761404)

[Empower Chromatography Data System <https://www.waters.com/513188>](https://www.waters.com/513188)

720007199, March 2021

© 2021 Waters Corporation. All Rights Reserved.